



CALIBRATION CENTER SCOPE OF CALIBRATION

DOK. NO	KEK EK 2
YAYIN TARİHİ	02.02.2009
REV.NO/TARİH	10/15.01.2016
SS/TS	1 / 2

NOT ACCREDITATED SCOPE

Measured Quantity Instrument or Gauge	Range	Measurement Conditions	Calibration and Measurement Capability Expended Uncertainty (k=2) (\pm)	Explanations
PRECIPITATION Rain gauge (Tipping, weighting...)	10 mm/hour $\leq I \leq$ 300 mm/hour	Electronic balance and peristaltic Pump	1 %	I: Precipitation Intensity, mm/hour Comparison Method
RADIATION Global Radiation (Pyronameter)	0 Watt/m ² $\leq G \leq$ 700 Watt/m ²	Under constant light intensity	1 %	G: Global Radiation, Watt/m ² Comparison Method
ELECTRICAL AC Current AC Current measuring instruments, Current meter, Multimeter, Datalogger	29 μ A $\leq I \leq$ 100 μ A	45 Hz $\leq f \leq$ 1 KHz	1 μ A	I : AC Current, A Comparison Method
	100,09 μ A $\leq I \leq$ 1 mA	45 Hz $\leq f \leq$ 1 KHz	5 μ A	
	1,09 mA $\leq I \leq$ 10 mA	45 Hz $\leq f \leq$ 1 KHz	20 μ A	
	10,09 mA $\leq I \leq$ 100 mA	45 Hz $\leq f \leq$ 1 KHz	400 μ A	
	100,09 mA $\leq I \leq$ 1 A	45 Hz $\leq f \leq$ 1 KHz	15 mA	
	1,09 A $\leq I \leq$ 10 A	45 Hz $\leq f \leq$ 1 KHz	30 mA	
DC Current DC Current measuring instruments, Current meter, Multimeter, Datalogger	0 μ A $\leq I \leq$ 300 μ A		0,5 μ A	I : DC Current, A Comparison Method
	300,09 μ A $\leq I \leq$ 3,29 mA		1 μ A	
	3,3 mA $\leq I \leq$ 32,9 mA		10 μ A	
	33 mA $\leq I \leq$ 329,9 mA		100 μ A	
	330 mA $\leq I \leq$ 1 A		500 μ A	
	1,09 A $\leq I \leq$ 3 A		5 mA	
	3,09 A $\leq I \leq$ 10 A		10 mA	
AC Voltage AC Voltage measuring instruments, Voltmeter, Multimeter, Datalogger	1 mV $\leq U \leq$ 30 mV	45 Hz $\leq f \leq$ 1 KHz	50 μ V	U : AC Voltage, V Comparison Method
	30,09 mV $\leq U \leq$ 100 mV	45 Hz $\leq f \leq$ 1 KHz	100 μ V	
	100,09 mV $\leq U \leq$ 1 V	45 Hz $\leq f \leq$ 1 KHz	600 μ V	
	1,09 V $\leq U \leq$ 10 V	45 Hz $\leq f \leq$ 1 KHz	10 mV	
	10,09 V $\leq U \leq$ 100 V	45 Hz $\leq f \leq$ 1 KHz	50 mV	
	100,09 V $\leq U \leq$ 650 V	45 Hz $\leq f \leq$ 1 KHz	400 mV	



Calibration
Center

CALIBRATION CENTER SCOPE OF CALIBRATION

DOK. NO	KEK EK 2
YAYIN TARİHİ	02.02.2009
REV.NO/TARİH	10/15.01.2016
SS/TS	2 / 2

Measured Quantity Instrument or Gauge	Range	Measurement Conditions	Calibration and Measurement Capability Expended Uncertainty ($k=2$) (\pm)	Explanations
DC Voltage DC Voltage measuring instruments, Voltmeter, Multimeter, Datalogger	$0 \leq U \leq 10 \text{ mV}$		$3 \mu\text{V}$	$U : \text{DC Voltage, V}$ Comparison Method
	$10,09 \text{ mV} \leq U \leq 329,9 \text{ mV}$		$15 \mu\text{V}$	
	$330 \text{ mV} \leq U \leq 1 \text{ V}$		$25 \mu\text{V}$	
	$1,09 \text{ mV} \leq U \leq 3,29 \text{ V}$		$100 \mu\text{V}$	
	$3,3 \text{ V} \leq U \leq 50 \text{ V}$		5 mV	
	$50,09 \text{ V} \leq U \leq 329,9 \text{ V}$		10 mV	
	$330 \text{ V} \leq U \leq 900 \text{ V}$		50 mV	
DC Resistance DC Resistance measuring instruments, Ohmmeter, Multimeter Datalogger	$0 \Omega \leq R \leq 32,99 \Omega$		$4 \text{ m}\Omega$	$R : \text{DC Resistance, } \Omega$ Comparison Method
	$33 \Omega \leq R \leq 329,9 \Omega$		$17 \text{ m}\Omega$	
	$330 \Omega \leq R \leq 1,09 \text{ K}\Omega$		$30 \text{ m}\Omega$	
	$1,1 \text{ K}\Omega \leq R \leq 109,99 \text{ K}\Omega$		5Ω	
	$110 \text{ K}\Omega \leq R \leq 1,09 \text{ M}\Omega$		60Ω	
	$1,1 \text{ M}\Omega \leq R \leq 10,99 \text{ M}\Omega$		$7 \text{ K}\Omega$	
	$11 \text{ M}\Omega \leq R \leq 109,99 \text{ M}\Omega$		$110 \text{ K}\Omega$	
Frequency Frequency measuring instruments, Frequency meter, Multimeter Datalogger	$0,01 \text{ Hz} \leq f \leq 119,99 \text{ Hz}$		8 mHz	$f : \text{Frequency, Hz}$ Comparison Method
	$120 \text{ Hz} \leq f \leq 1199,9 \text{ Hz}$		10 mHz	
	$1,2 \text{ KHz} \leq f \leq 119,9 \text{ KHz}$		1 Hz	
	$120 \text{ KHz} \leq f \leq 500 \text{ KHz}$		3 Hz	
ANGLE AIR DIRECTION Wind direction measuring instruments	$0^\circ \leq RY \leq 360^\circ$		4°	$RY : \text{Air (Wind) Direction, } {}^\circ$ (Degree) Comparison Method